



CERTIFICATE OF ACCREDITATION

ANSI National Accreditation Board
11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

X-Ray Now
2030 E. 4th Street, Suite D-110
Santa Ana, CA 92705

has been assessed by ANAB and meets the requirements of international standard

ISO/IEC 17025:2017

and national standard

ANSI/NCSL Z540-1-1994 (R2002)

while demonstrating technical competence in the field of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of activities to which this accreditation applies

AC-2672
Certificate Number


ANAB Approval

Certificate Valid Through: 04/17/2021
Version No. 001 Issued: 04/17/2019



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



ANSI National Accreditation Board

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 AND ANSI/NCSL Z540-1-1994 (R2002)

X-Ray Now

2030 E. 4th Street, Suite D-110
Santa Ana, CA 92705
Philip Abbott (714) 396-2519

CALIBRATION

Valid to: April 17, 2021

Certificate Number: AC-2672

Length – Dimensional Metrology

Table with 4 columns: Parameter / Equipment, Range, Expanded Uncertainty of Measurement (+/-), Reference Standard, Method and/or Equipment. Row 1: Coating Thickness Measuring Instruments by X-Ray Florescence Spectroscopy, (15 to 195) μin (0.38 μm to 4.9 μm), 0.56 μin + 4.8 % of reading, ASTM B568 (X-Ray Florescence)

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (k=2), corresponding to a confidence level of approximately 95%.

Notes:

- 1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2672.

Handwritten signature of R. D. P. over a horizontal line, with the title Vice President below it.

